

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/17/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

KMS
21-4/16/18

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:27
INTERNAL STANDARD	VERIFIED	11:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:28
BLANK TEST	0.00	11:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:30
BLANK TEST	0.00	11:31

