

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 09/24/2024

*RM*  
*211-1/2816*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:52
INTERNAL STANDARD	VERIFIED	11:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:54
BLANK TEST	0.00	11:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:55
BLANK TEST	0.00	11:56

