

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 10/01/2024

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24270

*RMS*  
*24-4/3896*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:21
INTERNAL STANDARD	VERIFIED	12:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:21
BLANK TEST	0.00	12:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:24
BLANK TEST	0.00	12:24

