

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/08/2024

RMS
24-44962

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36882
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:31
INTERNAL STANDARD	VERIFIED	11:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:34
BLANK TEST	0.00	11:35

