

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/17/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36882
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24279

RM
21-46857

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:52
INTERNAL STANDARD	VERIFIED	10:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:55

