

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 10/22/2024

*2MS*  
*24-46998*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:34
INTERNAL STANDARD	VERIFIED	08:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:34
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:37

