

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/12/2024

*RMS
2.1-50474*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:24
INTERNAL STANDARD	VERIFIED	14:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:24
BLANK TEST	0.00	14:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:26
BLANK TEST	0.00	14:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:27
BLANK TEST	0.00	14:28

