

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/19/2024

*RMS
24-5/16/18*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:23
INTERNAL STANDARD	VERIFIED	10:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:24
BLANK TEST	0.00	10:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:25
BLANK TEST	0.00	10:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:26
BLANK TEST	0.00	10:27

