

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/26/2024

RMS
21-52186

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:34
BLANK TEST	0.00	14:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:36
BLANK TEST	0.00	14:36

