

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

12MS
21-53285

Date: 12/03/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:05
INTERNAL STANDARD	VERIFIED	11:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:06
BLANK TEST	0.00	11:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:07
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:08
BLANK TEST	0.00	11:09

