

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 12/17/2024

*RMJ*  
*29-55267*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:49
INTERNAL STANDARD	VERIFIED	13:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:50
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:52
BLANK TEST	0.00	13:52

