

Rms 24-49730

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 11/04/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:54
INTERNAL STANDARD	VERIFIED	12:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:56
BLANK TEST	0.00	12:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:57
BLANK TEST	0.00	12:58

