

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 12/31/2024

*RM*  
*25-211*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:30
INTERNAL STANDARD	VERIFIED	11:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:31
BLANK TEST	0.00	11:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34

