

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 01/16/2024

LMs
24-2325

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:19
INTERNAL STANDARD	VERIFIED	14:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:20
BLANK TEST	0.00	14:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:21
BLANK TEST	0.00	14:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:22
BLANK TEST	0.00	14:23

