

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
24-23208

Date: 05/28/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:08
INTERNAL STANDARD	VERIFIED	12:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:08
BLANK TEST	0.00	12:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:10
BLANK TEST	0.00	12:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:11
BLANK TEST	0.00	12:11

