

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*RMS*  
*24 - 28/46*

Date: 06/25/2024

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36882  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:15
INTERNAL STANDARD	VERIFIED	11:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:15
BLANK TEST	0.00	11:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:18
BLANK TEST	0.00	11:18

