

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*EMS*  
*21-29537*

Date: 07/02/2024

OPERATOR NAME:  
 MICHAEL J. CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:52
INTERNAL STANDARD	VERIFIED	13:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:54
BLANK TEST	0.00	13:55

