

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/30/2024

RM
24-33707

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:39
INTERNAL STANDARD	VERIFIED	10:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:40
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:41
BLANK TEST	0.00	10:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:42
BLANK TEST	0.00	10:43

