

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM1
24-24/265

Date: 08/02/2024

OPERATOR NAME:
 MICHAEL J. CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:18
INTERNAL STANDARD	VERIFIED	10:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:20
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:21
BLANK TEST	0.00	10:21

