

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 08/20/2024

*RMS*  
*211-37163*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:45
INTERNAL STANDARD	VERIFIED	11:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:46
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:47
BLANK TEST	0.00	11:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:48
BLANK TEST	0.00	11:49

