

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dnt: 107308

Date: 08/27/2024

*RMS
211-38156*

OPERATOR NAME:
MICHAEL J CRIVE
Permit #: 36632
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 12:21 |
| INTERNAL STANDARD | VERIFIED | 12:21 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 12:22 |
| BLANK TEST | 0.00 | 12:23 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 12:23 |
| BLANK TEST | 0.00 | 12:24 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 12:24 |
| BLANK TEST | 0.00 | 12:25 |

