

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMJ
24-37163

Date: 09/02/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.08	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.08	10:36
BLANK TEST	0.00	10:37

