

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/03/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

*RMS
21-3930/8*

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:13
INTERNAL STANDARD	VERIFIED	10:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.08	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.08	10:15
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.08	10:16
BLANK TEST	0.00	10:17

