

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*RMS*

Date: 09/10/2024

*29-10473*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:57
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	12:58
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	12:59
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	13:00
BLANK TEST	0.00	13:01

