

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 09/17/2024

RMS
21-4/6/18

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:34
INTERNAL STANDARD	VERIFIED	11:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	11:35
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	11:36
BLANK TEST	0.00	11:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.06	11:37
BLANK TEST	0.00	11:38

