

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/24/2024

RM>
24-1/28/16

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:08
INTERNAL STANDARD	VERIFIED	12:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	12:09
BLANK TEST	0.00	12:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:10
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:12

