



SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM5

Date: 10/08/2024

21- 1/196Z

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:58
INTERNAL STANDARD	VERIFIED	11:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:59
BLANK TEST	0.00	12:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:00
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:01
BLANK TEST	0.00	12:02

