

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 10/15/2024

*RM3
21-1/68 5-11*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:15
INTERNAL STANDARD	VERIFIED	15:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:16
BLANK TEST	0.00	15:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:17
BLANK TEST	0.00	15:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:18
BLANK TEST	0.00	15:19

