



SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 10/22/2024

RM
24-1/6998

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:56
INTERNAL STANDARD	VERIFIED	08:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:57
BLANK TEST	0.00	08:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:59
BLANK TEST	0.00	09:00

