

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 10/29/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24270

RMJ
21-1/8/21

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:20
INTERNAL STANDARD	VERIFIED	10:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:21
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:22
BLANK TEST	0.00	10:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:23
BLANK TEST	0.00	10:24

