

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 11/12/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:42
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:44
BLANK TEST	0.00	13:45

RM7
21-50471
11/12/2024

