

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/02/2024

RMJ
2/1-53/60

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:10
INTERNAL STANDARD	VERIFIED	10:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:12
BLANK TEST	0.00	10:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:13
BLANK TEST	0.00	10:14

