

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/03/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

*RMS
24-53285*

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:16
INTERNAL STANDARD	VERIFIED	11:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:17
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:18
BLANK TEST	0.00	11:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:19
BLANK TEST	0.00	11:20

