

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RMS
24-55267

Date: 12/17/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:48
INTERNAL STANDARD	VERIFIED	13:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:49
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:50
BLANK TEST	0.00	13:51

