

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 01/07/2025

RMS
25-1092

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:09
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:12

