

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 01/14/2025

*Rm*  
*25-2027*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 35582  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35

