

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 01/21/2025

*(RM)*  
*25-3130*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:19
INTERNAL STANDARD	VERIFIED	13:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:19
BLANK TEST	0.00	13:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:20
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:21
BLANK TEST	0.00	13:22

