

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 01/28/2025

*RMS
25-1/395*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:25
INTERNAL STANDARD	VERIFIED	10:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:26
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:27
BLANK TEST	0.00	10:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:28
BLANK TEST	0.00	10:29

