

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
25-5217

Date: 02/03/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:38
INTERNAL STANDARD	VERIFIED	08:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:39
BLANK TEST	0.00	08:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:40
BLANK TEST	0.00	08:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:41
BLANK TEST	0.00	08:42

