

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RM
25-674/1

Date: 02/11/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:46
INTERNAL STANDARD	VERIFIED	09:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:47
BLANK TEST	0.00	09:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50

