

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/03/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 1271

RMS
75-9393

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:42
INTERNAL STANDARD	VERIFIED	11:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:42
BLANK TEST	0.00	11:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:44
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:45
BLANK TEST	0.00	11:45

