

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/18/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

RMS
25-11879

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:36
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:37
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39

