

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:09
INTERNAL STANDARD	VERIFIED	10:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:12
BLANK TEST	0.00	10:12

RMS
25-14055

