

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/08/2025

RMS
25-15/27

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator/Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:47
INTERNAL STANDARD	VERIFIED	09:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:50
BLANK TEST	0.00	09:51

