

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/25/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:30
BLANK TEST	0.00	09:31

RMS
25-28581

