

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:23
INTERNAL STANDARD	VERIFIED	09:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27

