

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/15/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:34
INTERNAL STANDARD	VERIFIED	10:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37

