

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/22/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:48
INTERNAL STANDARD	VERIFIED	07:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:48
BLANK TEST	0.00	07:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:49
BLANK TEST	0.00	07:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:50
BLANK TEST	0.00	07:51

