

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/29/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:35
INTERNAL STANDARD	VERIFIED	09:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:35
BLANK TEST	0.00	09:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:36
BLANK TEST	0.00	09:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:37
BLANK TEST	0.00	09:38

RMS
25-31/318

