

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:03
INTERNAL STANDARD	VERIFIED	10:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:04
BLANK TEST	0.00	10:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:05
BLANK TEST	0.00	10:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:06
BLANK TEST	0.00	10:07

